

Bonaduz, December 15, 2021

## Product Security Advisory – PSA-2021-03

### 1. Introduction

During our regular vulnerability monitoring procedures, we became aware of an issue with the Java library “log4j”. The vulnerability is referred to as CVE-2021-44228.

### 2. Affected Products

Hamilton Medical’s medical devices are not affected by the identified vulnerabilities, as none of our devices uses the Java library “log4j” or any Java-based applications.

### 3. Vulnerability Description

For a detailed description of the identified vulnerability, refer to CVE-2021-44228, published online at: <https://cve.mitre.org/cgi-bin/cvename.cgi?name=CVE-2021-44228>

### 4. Impact

As Hamilton Medical’s medical devices do not use the Java library “log4j”, the vulnerability CVE-2021-44228 has **no** impact on Hamilton Medical’s mechanical ventilators, respiratory humidifiers, cuff controllers, or medical accessories and consumables.

### 5. Mitigation

No customer interventions are necessary. In general, Hamilton Medical recommends:

- Prevent physical access to devices by unauthorized personnel
- Pay attention to notifications, alarms, and alerts
- Use only software from official Hamilton Medical channels and have it installed by an authorized technician

### 6. Contact Information

For further questions regarding the impact or mitigation of the vulnerabilities, contact your local sales or service manager. Alternatively, you can directly contact Hamilton Medical at:

[product-security.med.global@hamilton-medical.com](mailto:product-security.med.global@hamilton-medical.com)

## Revisions

Revision	Publication date	Updates
1.0	2021-12-15	Initial version

## Released by

<b>Lorenz Stähle</b> Group Team Leader Digital Solutions	<b>Annemarie Höft</b> Team Leader Regulatory Affairs